




EU/UKCA DECLARATION OF CONFORMITY

Under the sole responsibility of ControlAir LLC,
the following model(s) were found to be in compliance
with the essential requirements of the
UKCA Electromagnetic Compatibility Regulations 2016 & the
EMC Directive per EU Council Directive 2014/30/EU.

Models Certified: **Type 590X & Type 595XP
Pressure Transducers**

Relevant Standards: Immunity: EN 61326-1: 2013 Table 2
Emissions: EN 61326-1: 2013 Clause 7.2, Class A



Kevin J. McInnis
Director of Engineering

April 23, 2023
Date

Reference:

All equipment calibrated in accordance with ISO/IEC 17025
Test Report R-6798H-3